

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

February XX, 2026

Dear Distributor,

Records indicate that you have ordered or received product subject to this voluntary product recall (removal) as indicated in Attachment 1.

PLEASE DISTRIBUTE THIS INFORMATION TO ALL STAFF WITHIN YOUR FACILITY

Purpose of This Letter

B. Braun Medical Inc. (BBMI) is the manufacturer of the 21 Gauge Winged Infusion Set, an accessory included with MENTOR™ CPX™ 4 Tissue Expanders. BBMI informed Mentor of the recall (removal) of specific lots of the 21 Gauge Winged Infusion Set that our records indicate you received:

- A Winged Infusion Set included with CPX 4 Tissue Expanders (Reference Attachment 2).

Reason for the Voluntary Removal

BBMI has initiated a recall (removal) of specific lots of the 21 Gauge Winged Infusion Set due to the potential for the needle tip to be dull/blunt, difficult to advance, and/or break. Refer to the Action Required section of this communication for further instructions.

Risk to Health

BBMI's Risk to Health assessment states:

"To date there have been no reports of serious injury or death associated with this issue. There have been forty-one (41) cases of pain during insertion. If the product is used with a dull needle tip, it may not allow the septum of the tissue expander to be punctured which inhibits the procedure from moving forward. Tissue damage or damage to the septum may occur causing severe injury and/or pain and may lead to a delay in the procedure. If the product is used and the needle breaks, tissue damage or damage to the septum may occur causing injury and/or pain and may lead to delay or require intervention."

ACTION REQUIRED

1. Examine your inventory immediately to determine if you have product subject to this recall on hand and quarantine such product(s). If you have product subject to this recall, please maintain a copy of this notice with the quarantined product and keep a copy for your records.
2. Communicate the issue to relevant recall personnel or materials management personnel, or anyone else in your facility who needs to be informed. If any product subject to this recall has been forwarded to another facility, contact that facility to arrange return. Please consider including a copy of this recall letter when communicating.
3. Complete the Business Reply Form (BRF) (Attachment 3) confirming receipt of this notice and fax or email to [Enter Affiliate Information] within three (3) business days. **Please return the BRF even if you do not have product subject to this recall.**

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

4. Distributor are required to immediately return all impacted inventory (Attachment 1). Distributor must return product subject to this recall no later than **May 31, 2026** to **[Enter Affiliate Information]**. A credit will be issued for the **returned product that is received by May 31, 2026**.
5. Distributors are required to notify impacted customers using Attachment 4 (Direct Customer Letter).
6. Keep this notice visibly posted for awareness until all product subject to this recall has been returned to **[Enter Affiliate Information]**.

If you require any assistance with returning product, **please contact [Enter Affiliate Information] at [Enter Affiliate Information]**.

Other Information

At Mentor, our first priority is to our customers and their patients, and that includes the safe and effective use of our products. We recognize the recall of this product may be disruptive to your facility and we appreciate your assistance in this matter.

If you have additional questions regarding this communication or require any assistance with returning product, please contact **[Enter Affiliate Information]**.

As with any medical device, adverse reactions or quality problems experienced with the use of this product should be reported to your Sales Representative, directly to Mentor, or your National Health Authority. If you have any further questions related to this notice or if you need any additional communications, please contact your local Sales Representative.

The FSCA has been communicated to the relevant health authorities in the European Union, Switzerland, and the United Kingdom.

ATTACHMENTS:

- Attachment 1: Impacted Product Codes and Lots
- Attachment 2: Product Identification Tool
- Attachment 3: Business Reply Form (BRF)
- Attachment 4: Direct Customer Letter

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 1: Impacted Product Codes and Lots

EFFECTIVE IMMEDIATELY – DO NOT USE OR DISTRIBUTE THE FOLLOWING PRODUCT LOTS UNTIL THE 21 GAUGE INFUSION SET HAS BEEN REMOVED. REFER TO ACTION REQUIRED FOR FURTHER INSTRUCTIONS.

| PRODUCT DESCRIPTION | PRODUCT CODE | PRODUCT LOTS | | UDI |
|---|--------------|--------------|---------|----------------|
| CPX™ 4 Tissue Expander, Low Height, 250CC | 3548121 | 2153649 | | 00081317009061 |
| CPX™ 4 Tissue Expander, Low Height, 350CC | 3548122 | 2162733 | 2162614 | 00081317009078 |
| CPX™ 4 Tissue Expander, Low Height, 450CC | 3548123 | 2160712 | | 00081317009085 |
| CPX™ 4 Tissue Expander, Low Height, 550CC | 3548124 | 2159312 | | 00081317009092 |
| CPX™ 4 Tissue Expander, Low Height, 750CC | 3548126 | 2164676 | | 00081317009115 |
| CPX™ 4 Tissue Expander, Medium Height, 275CC | 3548221 | 2160708 | 2152570 | 00081317009122 |
| | | 2165160 | 2154201 | |
| CPX™ 4 Tissue Expander, Medium Height, 350CC | 3548222 | 2160707 | 2151903 | 00081317009139 |
| | | 2157852 | 2162509 | |
| | | 2164933 | 2159419 | |
| CPX™ 4 Tissue Expander, Medium Height, 450CC | 3548223 | 2159210 | 2159476 | 00081317009146 |
| | | 2164943 | 2162722 | |
| | | 2162116 | | |
| CPX™ 4 Tissue Expander, Medium Height, 550CC | 3548224 | 2151771 | 2164875 | 00081317009153 |
| | | 2162506 | 2159214 | |
| | | 2160711 | | |
| CPX™ 4 Tissue Expander, Tall Height, 450CC | 3548323 | 2159422 | 2164947 | 00081317009207 |
| CPX™ 4 Tissue Expander, Tall Height, 650CC | 3548325 | 2164949 | | 00081317009221 |
| CPX™ 4 Tissue Expander, Tall Height, 850CC | 3548327 | 2162354 | | 00081317009245 |
| CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 350CC | 3549122 | 2162191 | | 00081317009269 |
| CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 450CC | 3549123 | 2164677 | | 00081317009276 |
| CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 350CC | 3549222 | 2157853 | 2159372 | 00081317009320 |
| | | 2164948 | 2149039 | |
| CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 450CC | 3549223 | 2158742 | 2159371 | 00081317009337 |
| | | 2151503 | 2151793 | |
| CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 550CC | 3549224 | 2151564 | 2164873 | 00081317009344 |
| | | 2162037 | | |
| CPX™ 4 with Suture Tabs Tissue Expander, Tall Height, 250CC | 3549321 | 2164678 | | 00081317009375 |

FSCA# FA-0000131

Voluntary Field Safety Corrective Action (Product Recall/Removal) of 21 Gauge Winged Infusion Sets from Specific Lots of CPX™ 4 Tissue Expanders and Specific Lots of 21 Gauge Winged Infusion Set (B. Braun Medical 21 G X4.4CM Winged INF)

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 2: Product Identification Tool

Please refer to the pictures below for representative samples on how to identify the location of the subject product code and lots for impacted products by using the packaging labels, and for an example of where to locate the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits.

Tissue Expander Sales Unit Box



URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 3: Business Reply Form (BRF)

Business Reply Form (BRF)

Your timely response to this recall notification is requested. Please complete this form and fax or email it to [Enter Affiliate Information] or e-mail the form to [Enter Affiliate Information] **within 3 business days, even if you do not have product subject to this recall to return.**

If you have product subject to this recall to return, please make a photocopy of your completed Business Reply Form and enclose with your return. Thank you for your cooperation.

[Account Name]

[Account Address]

| | |
|--|-------------------|
| Print Name of Person Completing Business Reply Form: | Telephone Number: |
| Account Number (number used to order J&J product): | Date: |
| Email Address: | |
| Reference PO for credit, if needed. | |
| Signed*: | |
| <i>*Your signature provides confirmation that you have received and understood this notification</i> | |
| <i>Your comments are welcome.</i> | |

Product Inventory – please check one

- We have NO inventory of product subject to this recall (removal).
- We have product subject to this recall (removal) and are returning the following products:
- Our customer have returned product subject to this recall (removal) and we certify that following products were destroyed onsite:

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

21 Gauge Winged Infusion Sets from Specific Lots of CPX 4 Tissue Expanders

| PRODUCT CODE | PRODUCT LOT | | | QUANTITY RETURNING OR DESTROYED (EACHES) | | |
|--------------|-------------|--|--|--|--|--|
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 4: Direct Customer Letter

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

February XX, 2026

Dear Operating Room Supervisors, Materials Management Personnel, and Chief of Surgery

Records indicate that you have ordered or received product subject to this voluntary product recall (removal) as indicated in Attachment 1.

PLEASE DISTRIBUTE THIS INFORMATION TO ALL STAFF WITHIN YOUR FACILITY WHO USE THE CPX™ 4 TISSUE EXPANDERS

Purpose of This Letter

B. Braun Medical Inc. (BBMI) is the manufacturer of the 21 Gauge Winged Infusion Set, an accessory included with MENTOR™ CPX™ 4 Tissue Expanders. BBMI informed Mentor of the recall (removal) of specific lots of the 21 Gauge Winged Infusion Set that our records indicate you received:

- A Winged Infusion Set included with CPX 4 Tissue Expanders (Reference Attachment 2). **Note: All other components of the Tissue Expanders remain safe for use, are not subject to this removal and do not need to be returned.**

Reason for the Voluntary Removal

BBMI has initiated a recall (removal) of specific lots of the 21 Gauge Winged Infusion Set due to the potential for the needle tip to be dull/blunt, difficult to advance, and/or break. The 21 Gauge Winged Infusion Set is the only impacted component in the CPX Tissue Expander. **All other components of the Tissue Expanders remain safe for use, are not subject to this removal and do not need to be returned.** Refer to the Action Required section of this communication for further instructions.

Note: Per the Tissue Expanders instruction for use (IFU), a 21 gauge standard needle may be used in place of the BBMI Winged Infusion Set.

Risk to Health

BBMI's Risk to Health assessment states:

“To date there have been no reports of serious injury or death associated with this issue. There have been forty-one (41) cases of pain during insertion. If the product is used with a dull needle tip, it may not allow the septum of the tissue expander to be punctured which inhibits the procedure from moving forward. Tissue damage or damage to the septum may occur causing severe injury and/or pain and may lead to a delay in the procedure. If the product is used and the needle breaks, tissue damage or damage to the septum may occur causing injury and/or pain and may lead to delay or require intervention.”

FSCA# FA-0000131

Voluntary Field Safety Corrective Action (Product Recall/Removal) of 21 Gauge Winged Infusion Sets from Specific Lots of CPX™ 4 Tissue Expanders and Specific Lots of 21 Gauge Winged Infusion Set (B. Braun Medical 21 G X4.4CM Winged INF)

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

ACTION REQUIRED

1. Examine your inventory immediately to determine if you have product subject to this recall on hand and quarantine such product(s). If you have product subject to this recall, please maintain a copy of this notice with the quarantined product and keep a copy for your records.
 - **Note: if you have Tissue Expanders on consignment, no action is necessary for that inventory. A sales representative will manage this inventory.**
2. Communicate the issue to relevant recall personnel, operating room or materials management personnel, or anyone else in your facility who needs to be informed. If any product subject to this recall has been forwarded to another facility, contact that facility to arrange return. Please consider including a copy of this recall letter when communicating.
3. Complete the Business Reply Form (BRF) (Attachment 3) confirming receipt of this notice and fax or email to **[Enter Affiliate Information]** within three (3) business days. **Please return the BRF even if you do not have product subject to this recall.**
4. Customers are required to immediately **return or destroy** all unused 21 Gauge Winged Infusion Set subject to this recall that are in inventory. Customers must return product subject to this recall no later than **May 31, 2026** to **[Enter Affiliate Information]**. The Winged Infusion Set from the Tissue Expander **is not eligible for credit** because the 21 Gauge Winged Infusion Set is included in the package **at no charge**.
 - Note: Product can be destroyed or disposed following local procedures and regulations.**
5. To return product subject to this recall:
 - **Remove only the 21 Gauge Winged Infusion Set from any CPX 4 Tissue Expander package as listed in Attachment 1, Refer to Attachment 2 for an example of where to find the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits. This removal applies only to the 21 Gauge Winged Infusion Set, and the other components in the CPX 4 Tissue Expanders are safe to use and should NOT be returned.**
6. Keep this notice visibly posted for awareness until all product subject to this recall has been returned to **[Enter Affiliate Information]**.

If you require any assistance with returning product, please contact **[Enter Affiliate Information]** at **[Enter Affiliate Information]**.

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Other Information

At Mentor, our first priority is to our customers and their patients, and that includes the safe and effective use of our products. We recognize the recall of this product may be disruptive to your facility and we appreciate your assistance in this matter.

If you have additional questions regarding this communication or require any assistance with returning product, please contact [\[Enter Affiliate Information\]](#).

As with any medical device, adverse reactions or quality problems experienced with the use of this product should be reported to your Sales Representative, directly to Mentor, or your National Health Authority. If you have any further questions related to this notice or if you need any additional communications, please contact your local Sales Representative.

The FSCA has been communicated to the relevant health authorities in the European Union, Switzerland, and the United Kingdom.

ATTACHMENTS:

Attachment 1: Impacted Product Codes and Lots

Attachment 2: Product Identification Tool

Attachment 3: Business Reply Form (BRF)

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 1: Impacted Product Codes and Lots

EFFECTIVE IMMEDIATELY – DO NOT USE OR DISTRIBUTE THE FOLLOWING PRODUCT LOTS UNTIL THE 21 GAUGE INFUSION SET HAS BEEN REMOVED. REFER TO ACTION REQUIRED FOR FURTHER INSTRUCTIONS.

| PRODUCT DESCRIPTION | PRODUCT CODE | PRODUCT LOTS | | UDI |
|---|--------------|--------------|---------|----------------|
| CPX™ 4 Tissue Expander, Low Height, 250CC | 3548121 | 2153649 | | 00081317009061 |
| CPX™ 4 Tissue Expander, Low Height, 350CC | 3548122 | 2162733 | 2162614 | 00081317009078 |
| CPX™ 4 Tissue Expander, Low Height, 450CC | 3548123 | 2160712 | | 00081317009085 |
| CPX™ 4 Tissue Expander, Low Height, 550CC | 3548124 | 2159312 | | 00081317009092 |
| CPX™ 4 Tissue Expander, Low Height, 750CC | 3548126 | 2164676 | | 00081317009115 |
| CPX™ 4 Tissue Expander, Medium Height, 275CC | 3548221 | 2160708 | 2152570 | 00081317009122 |
| | | 2165160 | 2154201 | |
| CPX™ 4 Tissue Expander, Medium Height, 350CC | 3548222 | 2160707 | 2151903 | 00081317009139 |
| | | 2157852 | 2162509 | |
| | | 2164933 | 2159419 | |
| CPX™ 4 Tissue Expander, Medium Height, 450CC | 3548223 | 2159210 | 2159476 | 00081317009146 |
| | | 2164943 | 2162722 | |
| | | 2162116 | | |
| CPX™ 4 Tissue Expander, Medium Height, 550CC | 3548224 | 2151771 | 2164875 | 00081317009153 |
| | | 2162506 | 2159214 | |
| | | 2160711 | | |
| CPX™ 4 Tissue Expander, Tall Height, 450CC | 3548323 | 2159422 | 2164947 | 00081317009207 |
| CPX™ 4 Tissue Expander, Tall Height, 650CC | 3548325 | 2164949 | | 00081317009221 |
| CPX™ 4 Tissue Expander, Tall Height, 850CC | 3548327 | 2162354 | | 00081317009245 |
| CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 350CC | 3549122 | 2162191 | | 00081317009269 |
| CPX™ 4 with Suture Tabs Tissue Expander, Low Height, 450CC | 3549123 | 2164677 | | 00081317009276 |
| CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 350CC | 3549222 | 2157853 | 2159372 | 00081317009320 |
| | | 2164948 | 2149039 | |
| CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 450CC | 3549223 | 2158742 | 2159371 | 00081317009337 |
| | | 2151503 | 2151793 | |
| CPX™ 4 with Suture Tabs Tissue Expander, Medium Height, 550CC | 3549224 | 2151564 | 2164873 | 00081317009344 |
| | | 2162037 | | |
| CPX™ 4 with Suture Tabs Tissue Expander, Tall Height, 250CC | 3549321 | 2164678 | | 00081317009375 |

FSCA# FA-0000131

Voluntary Field Safety Corrective Action (Product Recall/Removal) of 21 Gauge Winged Infusion Sets from Specific Lots of CPX™ 4 Tissue Expanders and Specific Lots of 21 Gauge Winged Infusion Set (B. Braun Medical 21 G X4.4CM Winged INF)

URGENT: FIELD SAFETY NOTICE


21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 2: Product Identification Tool

Please refer to the pictures below for representative samples on how to identify the location of the subject product code and lots for impacted products by using the packaging labels, and for an example of where to locate the 21 Gauge Winged Infusion Set in CPX 4 Tissue Expander Kits.

Tissue Expander Sales Unit Box



Tissue Expander Sales Unit Box (Open)



URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

Attachment 3: Business Reply Form (BRF)

Business Reply Form (BRF)

Your timely response to this recall notification is requested. Please complete this form and fax or email it to **[Enter Affiliate Information]** or e-mail the form to **[Enter Affiliate Information]** **within 3 business days, even if you do not have product subject to this recall to return.**

If you have product subject to this recall to return, please make a photocopy of your completed Business Reply Form and enclose with your return. Thank you for your cooperation.

[Account Name]

[Account Address]

| | |
|--|-------------------|
| Print Name of Person Completing Business Reply Form: | Telephone Number: |
| Account Number (number used to order J&J product): | Date: |
| Email Address: | |
| Reference PO for credit, if needed. | |
| Signed*: | |
| <small>*Your signature provides confirmation that you have received and understood this notification</small> | |
| Your comments are welcome. | |

Product Inventory – please check one

- We have NO inventory of product subject to this recall (removal).
- We have product subject to this recall (removal) and are returning the following products:
- We have product subject to this recall (removal) and we certify that following products were destroyed onsite:

Please complete the tables on the next page.

URGENT: FIELD SAFETY NOTICE

21 Gauge Winged Infusion Sets (B. Braun Medical 21 G X4.4CM Winged INF) from Specific Lots of CPX™ 4 Tissue Expanders

– Voluntary Field Safety Corrective Action (Product Recall/Removal) –

21 Gauge Winged Infusion Sets from Specific Lots of CPX 4 Tissue Expanders

| PRODUCT CODE | PRODUCT LOT | | | QUANTITY RETURNING OR DESTROYED(EACHES) | | |
|--------------|-------------|--|--|---|--|--|
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |